

51 Scott Street West St. Catherines, ON L2R 1E2 ON: 905-688-5800 #110 - 1803 - 91 Street SW Edmonton, AB T6X OW8 EDM: 587 758 8367



## CERTIFICATE OF CALIBRATION

Certificate	Number

EE2018-1018-022

Date of Certification:

18-Oct-18

Equipment:

PosiTector UTG C probe

Probe Serial Number:

347436

Unit condition found:

New

Testing Environment:

Temperature:

23° ± 5°

Humidity:

up to 95%

Test Method:

This Ultrasonic Thickness Gauge was tested according to Manufacturer's specifications using a Certified 1018 Steel, 4-Step, Inch Calibration Block s/n 37962 traceable to NIST.

Step Thickness, Inch	Min. Reading Allowed	Gauge Measurement - Inch	Max. Reading Allowed
0.250	0.249	0.251	0.251
0.500	0.499	0.499	0.501
0.750	0.749	0.750	0.751
1.000	0.999	1.001	1.001

This document certifies that using the specified test method, the instrument met Manufacturer's Specifications of:

.040 to 5.000"

± 0.001"

1 to 125 mm

<u>+</u> 0.03 mm

Calibration Interval is dependant upon usage, handling, and storage, as well as site Quality Standards.

Unit Tested by:

RES

Email: info@stone-tucker.com Website: www.stone-tucker.com